



CALIBRATION CERTIFICATE

CUSTOMER: 4 Star Ventures **DATE:** May 9, 2014
CERTIFICATE# 29739-30160 **SERIAL#** 00340
MANUFACTURER: Schwabs **P.O.#**
MODEL: 6" Joint Torque Assembly **CYLINDER:** 4.409 In²
TEST GAUGE# 004 **CYL. S/N#** 00340A
RATIO: 25" Handle

<u>CALIBRATION PRESSURE</u>	<u>GAUGE DIAL READING</u>	<u>ACTUAL DIAL READING</u>
psi	ft.lbs	ft.lbs
327	3000	3000
653	6000	6000
980	9000	9000
1306	12000	12000

TECHNICIAN: Tam Pham

COMMENTS:

WE HERBY CERTIFY THAT THE ABOVE INSTRUMENT WAS CALIBRATED ACCORDING TO MANUFACTURER'S SPECIFICATIONS WHICH ARE +/- 2% OF FULL SCALE. THE CALIBRATION STANDARD IS TRACEABLE TO THE CANADIAN NATIONAL BUREAU OF STANDARDS AND COMPLIES WITH CANADIAN STANDARDS ASSOCIATION CAN3-299-3-85 AND A.N.S.I. B40-1.

Kane Instrumentation Ltd.
2109 - 84 Avenue
Edmonton, Alberta T6P 1K1
(780) 416-5263 (780) 416-8309 fax