



CALIBRATION CERTIFICATE

CUSTOMER: 4 Star Ventures **DATE:** Apr. 15, 2014
CERTIFICATE# 29529-29990 **SERIAL#** 19559
MANUFACTURER: Martin Decker **P.O.#** n/a
MODEL: 6" Joint Torque Assy **CYLINDER:** 3.68 in^
TEST GAUGE# 004 **CYL. S/N#** 19559A
RATIO: 32" Handle

CALIBRATION PRESSURE

psi
510
1020
1530
2040

GAUGE DIAL READING

ft.lbs
5000
10000
15000
20000

ACTUAL DIAL READING

ft.lbs
5000
10000
15000
20000


TECHNICIAN: Ken Mark

COMMENTS:

WE HERBY CERTIFY THAT THE ABOVE INSTRUMENT WAS CALIBRATED ACCORDING TO MANUFACTURER'S SPECIFICATIONS WHICH ARE +/- 2% OF FULL SCALE. THE CALIBRATION STANDARD IS TRACEABLE TO THE CANADIAN NATIONAL BUREAU OF STANDARDS AND COMPLIES WITH CANADIAN STANDARDS ASSOCIATION CAN3-299-3-85 AND A.N.S.I. B40-1.

Kane Instrumentation Ltd.
2109 - 84 Avenue
Edmonton, Alberta T6P 1K1
(780) 416-5263 (780) 416-8309 fax