



## CALIBRATION CERTIFICATE

**CUSTOMER:** 4 Star Ventures **DATE:** Apr. 17, 2014  
**CERTIFICATE#** 29533-30046 **SERIAL#** 29533  
**MANUFACTURER:** Wika **P.O.#** n/a  
**MODEL:** 6" Joint Torque Assy **CYLINDER:** 8.0 in<sup>^</sup>  
**TEST GAUGE#** 004 **CYL. S/N#** n/a  
**RATIO:** 18" Handle

<u>CALIBRATION PRESSURE</u>	<u>GAUGE DIAL READING</u>	<u>ACTUAL DIAL READING</u>
psi	ft.lbs	ft.lbs
167	2000	2000
334	4000	4000
501	6000	6000
668	8000	8050

TECHNICIAN: Tam Pham

**COMMENTS:**

WE HERBY CERTIFY THAT THE ABOVE INSTRUMENT WAS CALIBRATED ACCORDING TO MANUFACTURER'S SPECIFICATIONS WHICH ARE +/- 2% OF FULL SCALE. THE CALIBRATION STANDARD IS TRACEABLE TO THE CANADIAN NATIONAL BUREAU OF STANDARDS AND COMPLIES WITH CANADIAN STANDARDS ASSOCIATION CAN3-299-3-85 AND A.N.S.I. B40-1.

**Kane Instrumentation Ltd.**  
2109 - 84 Avenue  
Edmonton, Alberta T6P 1K1  
(780) 416-5263 (780) 416-8309 fax