



CALIBRATION CERTIFICATE

CUSTOMER: 4 Star Ventures **DATE:** May 8, 2014
CERTIFICATE# 29737-30158 **SERIAL#** M10683
MANUFACTURER: Martherne **P.O.#**
MODEL: 6" Joint Torque Assembly **CYLINDER:** 4.08 In²
TEST GAUGE# 004 **CYL. S/N#** KT108
RATIO: 32" Handle

| <u>CALIBRATION PRESSURE</u> | <u>GAUGE DIAL READING</u> | <u>ACTUAL DIAL READING</u> |
|-----------------------------|---------------------------|----------------------------|
| psi | ft.lbs | ft.lbs |
| 345 | 3750 | 3750 |
| 689 | 7500 | 7500 |
| 1034 | 11250 | 11250 |
| 1378 | 15000 | 15100 |

TECHNICIAN: Tam Pham

COMMENTS:

WE HERBY CERTIFY THAT THE ABOVE INSTRUMENT WAS CALIBRATED ACCORDING TO MANUFACTURER'S SPECIFICATIONS WHICH ARE +/- 2% OF FULL SCALE. THE CALIBRATION STANDARD IS TRACEABLE TO THE CANADIAN NATIONAL BUREAU OF STANDARDS AND COMPLIES WITH CANADIAN STANDARDS ASSOCIATION CAN3-299-3-85 AND A.N.S.I. B40-1.

Kane Instrumentation Ltd.
2109 - 84 Avenue
Edmonton, Alberta T6P 1K1
(780) 416-5263 (780) 416-8309 fax